

Glenosphere and Humeral Bearing Compatibility		Trabecular Metal Reverse Standard Polyethylene Liner Size 36mm (+0mm, +3mm, +6mm)	Trabecular Metal Reverse 65° Retentive Polyethylene Liner Size 36mm (+0mm, +3mm, +6mm)	Trabecular Metal Reverse Vivacit-E HXPE Liner Size 36mm (+0mm, +3mm, +6mm)	Trabecular Metal Reverse 65° Retentive Vivacit-E HXPE Liner Size 36mm (+0mm, +3mm, +6mm)	Anatomical Shoulder Inverse/Reverse Humeral PE-Inlay Size 36mm (+0mm, +3mm, +6mm)	Comprehensive Reverse ArComXL 44-36 Standard Humeral Bearing (+0mm, +3mm)	Comprehensive Reverse ArComXL 44-36 Retentive +3mm Humeral Bearing	Comprehensive Reverse E1 44-36 Standard Humeral Bearing (+0mm, +3mm)	Comprehensive Reverse E1 44-36 Retentive +3 Humeral Bearing	Comprehensive Reverse Prolong Poly Standard Humeral Bearing (40-36 mm) (+0mm, +3mm)	Comprehensive Reverse Prolong Polyethylene Retentive +3 mm Humeral Bearing (40-36 mm)	Comprehensive Reverse Vivacit-E HXPE Standard Humeral Bearing (40-36 mm) (+0mm, +3mm)	Comprehensive Reverse Vivacit-E HXPE Retentive +3 mm Humeral Bearing (40-36 mm)	Trabecular Metal Reverse Standard Polyethylene Liner Size 40mm (+0mm, +3mm, +6mm)	Trabecular Metal Reverse 65° Retentive Polyethylene Liner Size 40 mm (+0mm, +3mm, +6mm)	Trabecular Metal Reverse Vivacit-E HXPE Liner Size 40mm (0mm,+3mm, +6mm)	Trabecular Metal Reverse Vivacit-E HXPE 65° Retentive Liner Size 40mm (+0mm, +3mm, +6mm)	Anatomical Shoulder Inverse/Reverse Humeral PE-Inlay Size 40mm (0mm, +3mm, +6mm)	Comprehensive Reverse ArComXL Standard Humeral Bearing (44-41 mm) (+0mm, +3mm)	Comprehensive Reverse ArComXL Retentive +3 mm Humeral Bearing (44-41 mm)	Comprehensive Reverse E1 Standard Humeral Bearing (44-41 mm) (+0mm, +3mm)	Comprehensive Reverse E1 Retentive +3 mm Humeral Bearing (44-41 mm)	Comprehensive Reverse Prolong Poly Standard Humeral Bearing (40-40 mm) (+0mm, +3mm)	Comprehensive Reverse Prolong Polyethylene Retentive +3 mm Humeral Bearing (40-40 mm)	Comprehensive Reverse Vivacit-E HXPE Standard Humeral Bearing (40-40 mm) (+0mm, +3mm)	TESS Humeral - Inlay 36mm (6, 8, 10, 12mm)	TESS Humeral - Inlay 41mm (6, 8, 10, 12mm)	Identty Vivacit-E HXPE Retentive +0 mm Humeral Bearing (40-36 mm)	Identty Vivacit-E HXPE Retentive +0 mm Humeral Bearing (40-40 mm)					
Anaverse™ Glenoid Glenosphere 36-0°		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓																					
Anaverse Glenoid Glenosphere 40-0°															✓	✓	✓	✓	✓																
Anaverse Glenoid Glenosphere 41-0°																																			
Anaverse Glenoid Glenosphere 36-15°		✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓	✓																					
Anaverse Glenoid Glenosphere 40-15°															✓	✓	✓	✓	✓																
Anaverse Glenoid Glenosphere 41-15°																																			

Indicates "Combination has either not been assessed or is not functionally compatible"
 ✓ Indicates "Combination has been assessed as functionally compatible by Zimmer Biomet"

Humeral Tray and Humeral Bearing Compatibility	
	Comprehensive Reverse Prolong Poly Standard Humeral Bearing (40-36 mm) (+0mm, +3mm)
	Comprehensive Reverse Prolong Polyethylene Retentive +3 mm Humeral Bearing (40-36 mm)
	Comprehensive Reverse Vivacit-E HXPE Standard Humeral Bearing (40-36 mm) (+0mm, +3mm)
	Comprehensive Reverse Vivacit-E HXPE Retentive +3 mm Humeral Bearing (40-36 mm)
	Comprehensive Reverse Prolong Poly Standard Humeral Bearing (40-40 mm) (+0mm, +3mm)
	Comprehensive Reverse Prolong Polyethylene Retentive +3 mm Humeral Bearing (40-40 mm)
	Comprehensive Reverse Vivacit-E HXPE Standard Humeral Bearing (40-40 mm) (+0mm, +3mm)
	Comprehensive Reverse Vivacit-E HXPE Retentive +3 mm Humeral Bearing (40-40 mm)
	Identity Vivacit-E HXPE Retentive +0 mm Humeral Bearing (40-36 mm)
	Identity Vivacit-E HXPE Retentive +0 mm Humeral Bearing (40-40 mm)
Comprehensive Mini Humeral Trays	✓
Identity Humeral Trays	✓

Indicates "Combination has either not been assessed or is not functionally compatible"
 ✓ Indicates "Combination has been assessed as functionally compatible by Zimmer Biomet"

Note: As with all Reverse Shoulder Systems, the theoretical technical range of glenohumeral motion induced by using either the smallest diameter Glenosphere and/or a Retentive Humeral Liner/Bearing, is less than can be achieved when using a larger Glenosphere and/or a standard (non-retentive) Humeral Liner/Bearing. The surgeon should confirm, using the provided trials, that the selection of implant sizes fulfil the functional requirements of the patient in regard of range of motion.

Note: Some devices may not be available in your region. Please consult with your Zimmer Biomet sales representative for further information.

This represents functional compatibility and does not necessarily indicate regulatory approval status for marketing of product pairs.