



Signature™ ONE Guides & Bone Models Package Insert

(Valid from JUN 2019; replaces all previous versions)

**(DE) Signature™ ONE
Führung und Knochenmodelle Packungsbeilage**
**(ES) Signature™ ONE
Guías y modelos óseo Prospecto**
**(FR) Signature™ ONE
Notice des outils et modèles osseux**
**(IT) Signature™ ONE
Foglio illustrativo Guide e modelli di osso**
**(NL) Signature™ ONE
Geleiders en botmodellen Bijlschrift**

**(EN) Signature™ ONE
Guides & Bone Models Package Insert**
(Valid from JUN 2019; replaces all previous versions)

**English
SIGNATURE™ ONE GUIDES & BONE MODELS Package Insert**
(valid from JUN 2019; replaces all previous versions)

Before using this product, the operating surgeon should study carefully the following recommendations, warnings and instructions, as well as the available product-specific information (e.g., the applicable User's Guide, product literature, written Surgical Technique).

Zimmer Biomet is not liable for complications arising from the use of the device outside of its indicated uses, surgical technique, and similar matters outside the control of Zimmer Biomet.

1) Device Description
The Signature™ ONE System is composed of the following main components: the hardware (Signature™ ONE Fingerring and Bone Model) and the software (Internal manufacturing software applications & the Signature™ ONE Planer). The Signature™ ONE Guides and Bone Models are designed and manufactured of polyamide (nylon) using additive manufacturing (selective laser sintering), based on the approved and validated pre-approval process. The guides and bone models are provided non-sterile and sterilized at the hospital. They are used intra-operatively to assist the surgeon in reproducing the desired anatomical shape and to assist in the standard conventional surgical arthroplasty to allow conversion into a standard surgical technique at any time if needed during the operation.

The Signature™ ONE System uses the Zimmer Biomet Drive Portal (21 CFR 800.6310) for the interaction with external users (i.e., other surgeons and the Zimmer Biomet operators) using manufacturing software applications to prepare the patient cases for the surgeon.

Guide and Bone Model materials: polyamide and stainless steel

2) Indications for Use
The Signature™ ONE System is indicated, based on patient-specific radiological images with identifiable placement anatomical crosshairs, to assist the surgeon in the preparation of the shoulder arthroplasty, to assist in the surgical technique and to assist in the surgical procedures on patients not precluded from being radiologically scanned.

The Signature™ ONE System is to be used with the general components of the following shoulder implant systems in accordance with their indications and contraindications: Zimmer® Trabecular Metal™ Reverse Plus Shoulder, Comprehensive® Total Shoulder System, Comprehensive® Reverse Shoulder System and Comprehensive® Reverse Augmented Basepates.

The Signature™ ONE Guides and Bone Models are intended for single use only.

3) Contraindications
The Signature™ ONE System should not be used when the patient has metallic devices implanted that could interfere with the CT scan safety. Additionally, the Signature™ ONE System should not be used in cases where native bone is absent, or where a custom bone augmentgraft will be used, on surfaces intended to mate with the Signature™ ONE Guides.

4) Precautions and Warnings
The following are general precautions and warnings related to the use of Signature™ ONE Guides and Bone Models from Zimmer Biomet, referred to controls and instructions: Zimmer Biomet One System instruments within this insert. Other specific warnings and precautions are included in the Surgical Technique or User's Guide.

• Signature™ ONE Instruments should only be used by trained and qualified surgeons. The surgeon should attend a training given by Zimmer Biomet or its distributor prior to use.

• Signature™ ONE Instruments are to be used in conjunction with the implant system per the related pre-approval planning. The Signature™ ONE Instruments are provided in accordance with its respective packaging and sterilization information published by the implant manufacturer.

• Signature™ ONE Instruments are provided non-sterile. The instruments must be cleaned, packaged and sterilized before the use.

• Signature™ ONE Instruments are single use only and they must be discarded as biohazardous material after surgery.

• Signature™ ONE Instruments are not to be used for any other purpose than the intended use. Do not use for any other purpose than the intended use.

• Signature™ ONE Instruments are specific to the patient's anatomy as it was at the moment when the patient radiological images were acquired. Given the potential impact of patient movement on the radiological images, the surgeon has to assess the changes and in case of any doubt, the Signature™ ONE Instruments should not be used.

• Signature™ ONE Instruments should not be used after the expiration date as indicated on the package label.

• Signature™ ONE Instruments should not be used in conjunction with the Zimmer Biomet One System instruments.

• The user must verify that the Signature™ ONE Instruments are in good condition prior to the surgery, including that they are free of debris, cracks, or signs of deterioration.

• Maintenance and Sterilization should be handled carefully and they should not be placed under other heavy instruments or equipment to avoid deformation.

• Minimize excessive heat buildup due to friction between metallic instruments and adjacent surfaces, whether it is bone or other soft tissues. Excess heat buildup in instruments in contact with bone may cause heat necrosis and tissue deformation of the instruments.

• Signature™ ONE Instruments should not be used if the unique patient identifier (Signature™ Case ID) is not legible on the instrument.

• Signature™ ONE Instruments should be used in accordance with the instructions for use.

• Signature™ ONE Instruments that have been used and are being returned to Zimmer Biomet must be cleaned, packaged and sterilized prior to return, and be accompanied by a written confirmation of sterilization.

5) Cleaning
Signature™ ONE Instruments should not be cleaned with metal brushes or other cleaning equipment that can damage its surface. Use a soft bristle brush.

• Signature™ ONE Instruments must not come into contact with substances containing chlorine, fluorine, iodine, bromine, or other halogens. Instruments that are made entirely of plastic must not come into contact with strong acids or any, organic or anionic or non-fluorinating solvents, aromatic and/or halogen hydrocarbons or oxidizing chemicals.

The Signature™ ONE Instruments require manual cleaning or combination of cleaning and disinfection steps as follows for each instrument:

5.1) Manual Cleaning Instructions

1. Rinse the instrument under cold running tap water, for 2 minutes.

2. Prepare an enzymic detergent solution following the manufacturer's recommendation¹.

3. Immerse the instrument in the detergent solution and allow to soak for 10 minutes.

4. After the 10 minutes soak, use a soft bristle brush to brush the instrument, for 15 seconds, to remove all visible soil. Pay attention to crevices, mated surfaces and hard to reach areas. Use an appropriately sized lumen brush or channel brush to brush lumens and/or channels for 15 seconds.

5. Using the tap water jet, flush off to reach areas and closely mated surfaces (see areas labeled "B" in the images in the next column).

6. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

7. Immerse the instrument in the prepared enzymic solution and allow it to sonicate for 10 minutes.

8. Rinse the instrument under deionized water (DI) for 1 minute.

9. Dry the instrument using a clean lint-free cloth.

10. Inspect each instrument with the naked eye under normal lighting conditions to determine if all adherent visible soil (ex: blood protein substance and other debris) has been removed from surfaces, lumens, crevices and serrations, if applicable. If visible soil is seen, repeat the manual cleaning procedure.

5.2) Combination Cleaning and Disinfection Instructions

1. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

2. Prepare an enzymic detergent solution following the manufacturer's recommendations¹.

3. Immerse the instrument in the detergent solution and allow to soak for 10 minutes.

4. After the 10 minutes soak, use a soft bristle brush to brush the instrument, for 30 seconds, to remove all visible soil. Pay attention to crevices, mated surfaces and hard to reach areas. Use an appropriately sized lumen brush or channel brush to brush lumens and/or channels for 30 seconds.

5. Using the tap water jet, flush off to reach areas and closely mated surfaces (see areas labeled "B" in the images in the next column).

6. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

7. Immerse the instrument in the prepared enzymic solution and allow it to sonicate for 10 minutes.

8. Rinse the instrument under reverse osmosis-deionized water (RODI) for 1 minute.

9. Dry the instrument in a suitable washer/disinfectant basketed process through a standard instrument washer and/or inside our reprocessing cycle. The following parameters are essential for thorough cleaning and disinfection:

Table 1. Typical US Automated Washer/Disinfecter Cycle for Surgical Instruments

Step	Description
1	2 minutes pre-rinse with cold tap water
2	1 minute/20 seconds enzyme wash with hot tap water
3	15 seconds cold tap water rinse
4	2 minutes detergent wash with hot tap water (64°C/147°F)
5	2 minutes hot tap water rinse
6	15 seconds thermal rinse (90°C/176°F)

3) Contraindications
The Signature™ ONE System should not be used when the patient has metallic devices implanted that could interfere with the CT scan safety. Additionally, the Signature™ ONE System should not be used in cases where native bone is absent, or where a custom bone augmentgraft will be used, on surfaces intended to mate with the Signature™ ONE Guides.

4) Precautions and Warnings
The following are general precautions and warnings related to the use of Signature™ ONE Guides and Bone Models from Zimmer Biomet, referred to controls and instructions: Zimmer Biomet One System instruments within this insert. Other specific warnings and precautions are included in the Surgical Technique or User's Guide.

• Signature™ ONE Instruments should only be used by trained and qualified surgeons. The surgeon should attend a training given by Zimmer Biomet or its distributor prior to use.

• Signature™ ONE Instruments are to be used in conjunction with the implant system per the related pre-approval planning. The Signature™ ONE Instruments are provided in accordance with its respective packaging and sterilization information published by the implant manufacturer.

• Signature™ ONE Instruments are provided non-sterile. The instruments must be cleaned, packaged and sterilized before the use.

• Signature™ ONE Instruments are single use only and they must be discarded as biohazardous material after surgery.

• Signature™ ONE Instruments are not to be used for any other purpose than the intended use. Do not use for any other purpose than the intended use.

• Signature™ ONE Instruments are specific to the patient's anatomy as it was at the moment when the patient radiological images were acquired. Given the potential impact of patient movement on the radiological images, the surgeon has to assess the changes and in case of any doubt, the Signature™ ONE Instruments should not be used.

• Signature™ ONE Instruments should not be used after the expiration date as indicated on the package label.

• Signature™ ONE Instruments should not be used in conjunction with the Zimmer Biomet One System instruments.

• The user must verify that the Signature™ ONE Instruments are in good condition prior to the surgery, including that they are free of debris, cracks, or signs of deterioration.

• Maintenance and Sterilization should be handled carefully and they should not be placed under other heavy instruments or equipment to avoid deformation.

• Minimize excessive heat buildup due to friction between metallic instruments and adjacent surfaces, whether it is bone or other soft tissues. Excess heat buildup in instruments in contact with bone may cause heat necrosis and tissue deformation of the instruments.

• Signature™ ONE Instruments should not be used if the unique patient identifier (Signature™ Case ID) is not legible on the instrument.

• Signature™ ONE Instruments should be used in accordance with the instructions for use.

• Signature™ ONE Instruments that have been used and are being returned to Zimmer Biomet must be cleaned, packaged and sterilized prior to return, and be accompanied by a written confirmation of sterilization.

5) Cleaning
Signature™ ONE Instruments should not be cleaned with metal brushes or other cleaning equipment that can damage its surface. Use a soft bristle brush.

• Signature™ ONE Instruments must not come into contact with substances containing chlorine, fluorine, iodine, bromine, or other halogens. Instruments that are made entirely of plastic must not come into contact with strong acids or any, organic or anionic or non-fluorinating solvents, aromatic and/or halogen hydrocarbons or oxidizing chemicals.

The Signature™ ONE Instruments require manual cleaning or combination of cleaning and disinfection steps as follows for each instrument:

5.1) Manual Cleaning Instructions

1. Rinse the instrument under cold running tap water, for 2 minutes.

2. Prepare an enzymic detergent solution following the manufacturer's recommendation¹.

3. Immerse the instrument in the detergent solution and allow to soak for 10 minutes.

4. After the 10 minutes soak, use a soft bristle brush to brush the instrument, for 15 seconds, to remove all visible soil. Pay attention to crevices, mated surfaces and hard to reach areas. Use an appropriately sized lumen brush or channel brush to brush lumens and/or channels for 15 seconds.

5. Using the tap water jet, flush off to reach areas and closely mated surfaces (see areas labeled "B" in the images in the next column).

6. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

7. Immerse the instrument in the prepared enzymic solution and allow it to sonicate for 10 minutes.

8. Rinse the instrument under deionized water (DI) for 1 minute.

9. Dry the instrument using a clean lint-free cloth.

10. Inspect each instrument with the naked eye under normal lighting conditions to determine if all adherent visible soil (ex: blood protein substance and other debris) has been removed from surfaces, lumens, crevices and serrations, if applicable. If visible soil is seen, repeat the manual cleaning procedure.

5.2) Combination Cleaning and Disinfection Instructions

1. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

2. Prepare an enzymic detergent solution following the manufacturer's recommendations¹.

3. Immerse the instrument in the detergent solution and allow to soak for 10 minutes.

4. After the 10 minutes soak, use a soft bristle brush to brush the instrument, for 30 seconds, to remove all visible soil. Pay attention to crevices, mated surfaces and hard to reach areas. Use an appropriately sized lumen brush or channel brush to brush lumens and/or channels for 30 seconds.

5. Using the tap water jet, flush off to reach areas and closely mated surfaces (see areas labeled "B" in the images in the next column).

6. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

7. Immerse the instrument in the prepared enzymic solution and allow it to sonicate for 10 minutes.

8. Rinse the instrument under reverse osmosis-deionized water (RODI) for 1 minute.

9. Dry the instrument in a suitable washer/disinfectant basketed process through a standard instrument washer and/or inside our reprocessing cycle. The following parameters are essential for thorough cleaning and disinfection:

Table 1. Typical US Automated Washer/Disinfecter Cycle for Surgical Instruments

Step	Description
1	2 minutes pre-rinse with cold tap water
2	1 minute/20 seconds enzyme wash with hot tap water
3	15 seconds cold tap water rinse
4	2 minutes detergent wash with hot tap water (64°C/147°F)
5	2 minutes hot tap water rinse
6	15 seconds thermal rinse (90°C/176°F)

3) Contraindications
The Signature™ ONE System should not be used when the patient has metallic devices implanted that could interfere with the CT scan safety. Additionally, the Signature™ ONE System should not be used in cases where native bone is absent, or where a custom bone augmentgraft will be used, on surfaces intended to mate with the Signature™ ONE Guides.

4) Precautions and Warnings
The following are general precautions and warnings related to the use of Signature™ ONE Guides and Bone Models from Zimmer Biomet, referred to controls and instructions: Zimmer Biomet One System instruments within this insert. Other specific warnings and precautions are included in the Surgical Technique or User's Guide.

• Signature™ ONE Instruments should only be used by trained and qualified surgeons. The surgeon should attend a training given by Zimmer Biomet or its distributor prior to use.

• Signature™ ONE Instruments are to be used in conjunction with the implant system per the related pre-approval planning. The Signature™ ONE Instruments are provided in accordance with its respective packaging and sterilization information published by the implant manufacturer.

• Signature™ ONE Instruments are provided non-sterile. The instruments must be cleaned, packaged and sterilized before the use.

• Signature™ ONE Instruments are single use only and they must be discarded as biohazardous material after surgery.

• Signature™ ONE Instruments are not to be used for any other purpose than the intended use. Do not use for any other purpose than the intended use.

• Signature™ ONE Instruments are specific to the patient's anatomy as it was at the moment when the patient radiological images were acquired. Given the potential impact of patient movement on the radiological images, the surgeon has to assess the changes and in case of any doubt, the Signature™ ONE Instruments should not be used.

• Signature™ ONE Instruments should not be used after the expiration date as indicated on the package label.

• Signature™ ONE Instruments should not be used in conjunction with the Zimmer Biomet One System instruments.

• The user must verify that the Signature™ ONE Instruments are in good condition prior to the surgery, including that they are free of debris, cracks, or signs of deterioration.

• Maintenance and Sterilization should be handled carefully and they should not be placed under other heavy instruments or equipment to avoid deformation.

• Minimize excessive heat buildup due to friction between metallic instruments and adjacent surfaces, whether it is bone or other soft tissues. Excess heat buildup in instruments in contact with bone may cause heat necrosis and tissue deformation of the instruments.

• Signature™ ONE Instruments should not be used if the unique patient identifier (Signature™ Case ID) is not legible on the instrument.

• Signature™ ONE Instruments should be used in accordance with the instructions for use.

• Signature™ ONE Instruments that have been used and are being returned to Zimmer Biomet must be cleaned, packaged and sterilized prior to return, and be accompanied by a written confirmation of sterilization.

5) Cleaning
Signature™ ONE Instruments should not be cleaned with metal brushes or other cleaning equipment that can damage its surface. Use a soft bristle brush.

• Signature™ ONE Instruments must not come into contact with substances containing chlorine, fluorine, iodine, bromine, or other halogens. Instruments that are made entirely of plastic must not come into contact with strong acids or any, organic or anionic or non-fluorinating solvents, aromatic and/or halogen hydrocarbons or oxidizing chemicals.

The Signature™ ONE Instruments require manual cleaning or combination of cleaning and disinfection steps as follows for each instrument:

5.1) Manual Cleaning Instructions

1. Rinse the instrument under cold running tap water, for 2 minutes.

2. Prepare an enzymic detergent solution following the manufacturer's recommendation¹.

3. Immerse the instrument in the detergent solution and allow to soak for 10 minutes.

4. After the 10 minutes soak, use a soft bristle brush to brush the instrument, for 15 seconds, to remove all visible soil. Pay attention to crevices, mated surfaces and hard to reach areas. Use an appropriately sized lumen brush or channel brush to brush lumens and/or channels for 15 seconds.

5. Using the tap water jet, flush off to reach areas and closely mated surfaces (see areas labeled "B" in the images in the next column).

6. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

7. Immerse the instrument in the prepared enzymic solution and allow it to sonicate for 10 minutes.

8. Rinse the instrument under deionized water (DI) for 1 minute.

9. Dry the instrument using a clean lint-free cloth.

10. Inspect each instrument with the naked eye under normal lighting conditions to determine if all adherent visible soil (ex: blood protein substance and other debris) has been removed from surfaces, lumens, crevices and serrations, if applicable. If visible soil is seen, repeat the manual cleaning procedure.

5.2) Combination Cleaning and Disinfection Instructions

1. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

2. Prepare an enzymic detergent solution following the manufacturer's recommendations¹.

3. Immerse the instrument in the detergent solution and allow to soak for 10 minutes.

4. After the 10 minutes soak, use a soft bristle brush to brush the instrument, for 30 seconds, to remove all visible soil. Pay attention to crevices, mated surfaces and hard to reach areas. Use an appropriately sized lumen brush or channel brush to brush lumens and/or channels for 30 seconds.

5. Using the tap water jet, flush off to reach areas and closely mated surfaces (see areas labeled "B" in the images in the next column).

6. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

7. Immerse the instrument in the prepared enzymic solution and allow it to sonicate for 10 minutes.

8. Rinse the instrument under reverse osmosis-deionized water (RODI) for 1 minute.

9. Dry the instrument in a suitable washer/disinfectant basketed process through a standard instrument washer and/or inside our reprocessing cycle. The following parameters are essential for thorough cleaning and disinfection:

Table 1. Typical US Automated Washer/Disinfecter Cycle for Surgical Instruments

Step	Description
1	2 minutes pre-rinse with cold tap water
2	1 minute/20 seconds enzyme wash with hot tap water
3	15 seconds cold tap water rinse
4	2 minutes detergent wash with hot tap water (64°C/147°F)
5	2 minutes hot tap water rinse
6	15 seconds thermal rinse (90°C/176°F)

3) Contraindications
The Signature™ ONE System should not be used when the patient has metallic devices implanted that could interfere with the CT scan safety. Additionally, the Signature™ ONE System should not be used in cases where native bone is absent, or where a custom bone augmentgraft will be used, on surfaces intended to mate with the Signature™ ONE Guides.

4) Precautions and Warnings
The following are general precautions and warnings related to the use of Signature™ ONE Guides and Bone Models from Zimmer Biomet, referred to controls and instructions: Zimmer Biomet One System instruments within this insert. Other specific warnings and precautions are included in the Surgical Technique or User's Guide.

• Signature™ ONE Instruments should only be used by trained and qualified surgeons. The surgeon should attend a training given by Zimmer Biomet or its distributor prior to use.

• Signature™ ONE Instruments are to be used in conjunction with the implant system per the related pre-approval planning. The Signature™ ONE Instruments are provided in accordance with its respective packaging and sterilization information published by the implant manufacturer.

• Signature™ ONE Instruments are provided non-sterile. The instruments must be cleaned, packaged and sterilized before the use.

• Signature™ ONE Instruments are single use only and they must be discarded as biohazardous material after surgery.

• Signature™ ONE Instruments are not to be used for any other purpose than the intended use. Do not use for any other purpose than the intended use.

• Signature™ ONE Instruments are specific to the patient's anatomy as it was at the moment when the patient radiological images were acquired. Given the potential impact of patient movement on the radiological images, the surgeon has to assess the changes and in case of any doubt, the Signature™ ONE Instruments should not be used.

• Signature™ ONE Instruments should not be used after the expiration date as indicated on the package label.

• Signature™ ONE Instruments should not be used in conjunction with the Zimmer Biomet One System instruments.

• The user must verify that the Signature™ ONE Instruments are in good condition prior to the surgery, including that they are free of debris, cracks, or signs of deterioration.

• Maintenance and Sterilization should be handled carefully and they should not be placed under other heavy instruments or equipment to avoid deformation.

• Minimize excessive heat buildup due to friction between metallic instruments and adjacent surfaces, whether it is bone or other soft tissues. Excess heat buildup in instruments in contact with bone may cause heat necrosis and tissue deformation of the instruments.

• Signature™ ONE Instruments should not be used if the unique patient identifier (Signature™ Case ID) is not legible on the instrument.

• Signature™ ONE Instruments should be used in accordance with the instructions for use.

• Signature™ ONE Instruments that have been used and are being returned to Zimmer Biomet must be cleaned, packaged and sterilized prior to return, and be accompanied by a written confirmation of sterilization.

5) Cleaning
Signature™ ONE Instruments should not be cleaned with metal brushes or other cleaning equipment that can damage its surface. Use a soft bristle brush.

• Signature™ ONE Instruments must not come into contact with substances containing chlorine, fluorine, iodine, bromine, or other halogens. Instruments that are made entirely of plastic must not come into contact with strong acids or any, organic or anionic or non-fluorinating solvents, aromatic and/or halogen hydrocarbons or oxidizing chemicals.

The Signature™ ONE Instruments require manual cleaning or combination of cleaning and disinfection steps as follows for each instrument:

5.1) Manual Cleaning Instructions

1. Rinse the instrument under cold running tap water, for 2 minutes.

2. Prepare an enzymic detergent solution following the manufacturer's recommendation¹.

3. Immerse the instrument in the detergent solution and allow to soak for 10 minutes.

4. After the 10 minutes soak, use a soft bristle brush to brush the instrument, for 15 seconds, to remove all visible soil. Pay attention to crevices, mated surfaces and hard to reach areas. Use an appropriately sized lumen brush or channel brush to brush lumens and/or channels for 15 seconds.

5. Using the tap water jet, flush off to reach areas and closely mated surfaces (see areas labeled "B" in the images in the next column).

6. Prepare an enzymic detergent solution following the manufacturer's recommendation in an ultrasonic unit.

