

Omni™ Suite System Initial Release 1.0

Important Information for Users

- With manual workflow, if the workflow is advanced from anesthesia start to surgery start by mistake and then reset to anesthesia start, the anesthesia timer will not reset. If it is advanced from surgery start to surgery stop by mistake and then reset to surgery start the surgery timer will not be reset. Please refer to Omni™ Suite System User Manual (IFU) for **Manual Events** section and more details.